

Policy Number	OTH903.025
Policy Effective Date	07/01/2026

Eyelid Thermal Pulsation for the Treatment of Dry Eye Syndrome

Table of Contents
Coverage
Policy Guidelines
Description
Rationale
Coding
References
Policy History

Related Policies (if applicable)
None

Disclaimer

Carefully check state regulations and/or the member contract.

Each benefit plan, summary plan description or contract defines which services are covered, which services are excluded, and which services are subject to dollar caps or other limitations, conditions or exclusions. Members and their providers have the responsibility for consulting the member's benefit plan, summary plan description or contract to determine if there are any exclusions or other benefit limitations applicable to this service or supply. **If there is a discrepancy between a Medical Policy and a member's benefit plan, summary plan description or contract, the benefit plan, summary plan description or contract will govern.**

Coverage

Eyelid thermal pulsation therapy (e.g., LipiFlow® Thermal Pulsation System, TearCare® System) to treat dry eye syndrome **is considered experimental, investigational and/or unproven.**

Policy Guidelines

None.

Description

Thermal pulsation is a treatment option for meibomian gland dysfunction (MGD). Meibomian gland dysfunction is recognized as the major cause of dry eye syndrome. Thermal pulsation applies heat to the palpebral surfaces of the upper and lower eyelids directly over the meibomian glands, while simultaneously applying graded pulsatile pressure to the outer eyelid surfaces, thereby expressing the meibomian glands.

Dry Eye Syndrome

Dry eye syndrome, dry eye disease, or dysfunctional tear syndrome, either alone or in combination with other conditions, is a frequent cause of ocular irritation that leads patients to seek ophthalmologic care. It is estimated to affect between 5% and 50% of the population worldwide. (1) Based on data from 2013, an estimated 16.4 million Americans have dry eye syndrome. (2) The prevalence of dry eye syndrome increases with age, especially in postmenopausal women. For both sexes, prevalence is more than 3 times higher in individuals 50 years of age or older compared to those 18 to 49 years of age. Meibomian gland dysfunction is considered to be the most common cause of dry eye syndrome. (3)

In a 2022 meta-analysis of 3 United States studies, the prevalence of dry eye ranged from 5% to 14% with an estimated pooled prevalence of 8%. The prevalence of MGD ranged from 10% to 55%. Over a 5-year period, the incidence of dry eye was 3% among individuals aged 18 years and older, and 8% among those aged 68 years and older. (4) Prevention and treatment of dry eye syndrome are expected to be of greater importance as the population ages.

Treatment

Current treatment options for MGD include physical expression to relieve the obstruction, administration of heat (warm compresses) to the eyelids to liquefy solidified meibomian gland contents, eyelid scrubs to relieve external meibomian gland orifice blockage, and medications (e.g., antibiotics, topical corticosteroids) to mitigate infection and inflammation of the eyelids. (3, 5-7) These treatment options, however, have shown limited clinical efficacy, and often require a trial-and-error approach. For example, physical expression can be very painful given the amount of force needed to express obstructed glands. Warm compress therapy can be time-consuming and labor intensive, and there is limited evidence that medications relieve MGD. (6) While the symptoms of dry eye syndrome often improve with treatment, the disease usually is not curable and may lead to substantial patient and physician frustration. (3, 7) Dry eyes can be a cause of visual morbidity and may compromise results of corneal, cataract, and refractive surgery. Inadequate treatment of dry eye syndrome may result in increased ocular discomfort, blurred vision, reduced quality of life, and decreased productivity.

Regulatory Status

Eyelid thermal pulsation systems (FDA product code: ORZ) cleared by the U.S. Food and Drug Administration (FDA) are summarized in Table 1. Other office-based devices that deliver constant rather than pulsed heat to the eyelid (for example, MiBoFlow®/MiBo Thermoflo®, Blephasteam®) are not discussed in this policy.

Table 1. Eyelid Thermal Pulsation Systems Cleared by the FDA

Device	Manufacturer	Location	Original Date Cleared/ Approved	Original De Novo or 510(k) Number or PMA	Indication

LipiFlow® Thermal Pulsation System	TearScience	Morrisville, NC	2011*	DEN100017*	'For the application of localized heat and pressure therapy in adult patients with chronic cystic conditions of the eyelids, including meibomian gland dysfunction (MGD), also known as evaporative dry eye or lipid deficiency dry eye.'
iLux® System	Tear Film Innovations ^a	San Diego, CA	2017	K172645	'For the application of localized heat and pressure therapy in adult patients with chronic diseases of the eyelids, including MGD, also known as evaporative dry eye.'
Systane® iLux2®	Tear Film Innovations ^a	Carlsbad, CA	2020	K200400	'For the application of localized heat and pressure therapy in adult patients with MGD, which is associated with evaporative dry eye, and to capture/store digital images and video of the meibomian glands'
TearCare® System	Sight Sciences	Menlo Park, CA	2021	K213045	'For the application of localized heat and pressure therapy in adult patients with evaporative dry eye disease due to MGD, when used in conjunction with manual expression of the meibomian glands.'
TearCare® MGX™	Sight Sciences	Menlo Park, CA	2023	K231084	'For the application of localized heat therapy in adult patients with evaporative dry eye disease due to MGD, when used in conjunction with manual expression of the meibomian glands.'

PMA: premarket approval.

*Other 501(k) numbers are associated with more recent versions of the device.

^aAlcon, a division of Novartis, acquired Tear Film Innovations in 2018.

NOTE: This section is to be used for informational purposes only. FDA approval alone is not a basis for coverage.

Rationale

Medical policies assess clinical evidence to determine whether the use of a technology improves the net health outcome. Broadly defined, health outcomes are length of life, quality of life, and ability to function, including benefits and harms. Every clinical condition has specific outcomes that are important to patients and to managing the course of that condition. Validated outcome measures are necessary to ascertain whether a condition improves or worsens; and whether the magnitude of that change is clinically significant. The net health outcome is a balance of benefits and harms.

To assess whether the evidence is sufficient to draw conclusions about the net health outcome of a technology, 2 domains are examined: the relevance and the quality and credibility. To be relevant, studies must represent one or more intended clinical use of the technology in the intended population and compare an effective and appropriate alternative at a comparable intensity. For some conditions, the alternative will be supportive care or surveillance. The quality and credibility of the evidence depend on study design and conduct, minimizing bias and confounding that can generate incorrect findings. The randomized controlled trial (RCT) is preferred to assess efficacy; However, in some circumstances, nonrandomized studies may be adequate. RCTs are rarely large enough or long enough to capture less common adverse events and long-term effects. Other types of studies can be used for these purposes and to assess generalizability to broader clinical populations and settings of clinical practice.

Dry Eye Syndrome

Clinical Context and Therapy Purpose

The purpose of eyelid thermal pulsation in individuals who have dry eye syndrome is to provide a treatment option that is an alternative to or an improvement on existing therapies.

The following PICO was used to select literature to inform this policy.

Populations

The relevant population(s) of interest is individuals with dry eye syndrome. Dry eye syndrome is often classified into the aqueous-deficient subtype or the evaporative subtype, although classification is not mutually exclusive. Dry eye syndrome is a multifactorial disease of the ocular surface that may require a combination approach to treatment. Meibomian gland dysfunction (MGD), characterized by changes in gland secretion with or without concomitant gland obstruction, is recognized as the most common cause of evaporative dry eye and may also play a role in aqueous-deficient dry eye.

Interventions

The therapy being considered is eyelid thermal pulsation. The LipiFlow Thermal Pulsation System is one of the devices developed to relieve MGD. This device heats the palpebral surfaces of both the upper and lower eyelids, while applying graded pulsatile pressure to the outer eyelid surfaces. The LipiFlow System is composed of a disposable ocular component and a handheld control system. Following application of a topical anesthetic, the heated inner portion of the LipiFlow eyecup is applied to the conjunctival surface of the upper and lower eyelids. The outer portion of the device covers the skin surface of the upper and lower eyelids. The device massages the eyelids with cyclical pressure from the base of the meibomian glands in the direction of the gland orifices, thereby expressing the glands during heating. The TearCare System is another Food and Drug Administration (FDA)-cleared device intended for the application of localized heat therapy in adult patients with evaporative dry eye disease due to MGD, when used in conjunction with manual expression of the meibomian glands. The system uses 2 wearable devices affixed to the eyelids and a central controller to gradually raise the eyelid temperature to 45°C, melting gland obstructions. This thermal procedure is followed by manual gland expression using the provided purpose-designed device.

Comparators

The following practices are currently being used to treat dry eye syndrome: standard treatment with warm compresses and eyelid massage. Current treatment options for MGD include physical expression to relieve the obstruction, administration of heat (warm compresses) to the eyelids to liquefy solidified meibomian gland contents, eyelid scrubs to relieve external meibomian gland orifice blockage, and medications (e.g., antibiotics, topical corticosteroids) to mitigate infection and inflammation of the eyelids.

Outcomes

The general outcomes of interest are symptoms, morbid events, and functional outcomes.

Tear break-up time (TBUT) is measured in seconds. Practice parameters from the American Academy of Ophthalmology (2013) have indicated that a tear break-up time of <10 seconds is considered abnormal. (7)

The Ocular Surface Disease Index (OSDI) assesses the patient's frequency and severity of dry eye symptoms in specific contexts during the week prior to the examination. The minimal clinically important difference for the OSDI ranges from 4.5 to 7.3 for mild or moderate disease. The overall OSDI score defines the ocular surface as normal (0 to 12 points) or as having mild (13 to 22 points), moderate (23 to 32 points), or severe (33 to 100 points) disease. (8)

The Standard Patient Evaluation for Eye Dryness (SPEED) questionnaire is a self-reported measure of the frequency and severity of dryness, grittiness, scratchiness, soreness, irritation, burning, watering, and eye fatigue. It was developed by TearScience and validated in a 2013 study funded by TearScience. (9) In this validation study, the mean SPEED score of symptomatic subjects was 21.0 and the mean of asymptomatic subjects was 6.25.

The Meibomian Gland Expression Score (MGES) is a numerical rating used to evaluate the ease with which the meibomian glands in the eyelids can release oil (meibum). A higher score suggests more difficulty in oil expression, potentially indicating MGD. Typically, the score is determined by the number of glands that can be expressed without difficulty, where 0 indicates all glands express oil easily, and 3 indicates none of the glands express oil at all. The preferred approach is to record the sum of scores for each gland expressed, to achieve a composite score. If 8 glands are expressed, then the score range is 0 to $(8 \times 3) = 24$. (10)

Study Selection Criteria

Methodologically credible studies were selected using the following principles:

- To assess efficacy outcomes, comparative controlled prospective trials were sought, with a preference for RCTs.
- In the absence of such trials, comparative observational studies were sought, with a preference for prospective studies.
- To assess longer term outcomes and adverse events, single-arm studies that capture longer periods of follow-up and/or larger populations were sought.
- Studies with duplicative or overlapping populations were excluded.

Systematic Reviews

Ben Ephraim Noyman et al. (2025) conducted a systematic review of 45 studies that compared device-based MDG treatment to conservative therapy. (11) Of the 45 studies, 9 evaluated LipiFlow, 2 evaluated iLux, and 2 evaluated TearCare. For TBUT, none of the thermal pulsation devices had a significant effect compared to conservative treatment. Similarly, none of the MGD symptom scores were significantly different between thermal pulsation devices and conservative treatment.

In a 2024 Cochrane review, Pucker et al. evaluated the effectiveness of LipiFlow for treating dry eye disease and the safety of this treatment compared to sham and/or other treatments for MGD. (12) Across 13 RCTs, published through October 2022, a total of 1155 participants were randomized, with each study ranging from 28 to 236 participants. Of these trials, 6 took place in the USA, 3 in China, 2 in Thailand, one in France, and one in Italy. Eight trials were single-center, 4 were multicenter, and one did not specify the number of centers involved. The participants consisted of 66% females, ranging in age from 19 to 86 years. The LipiFlow treatment was assessed as a stand-alone intervention against basic warm compresses in 5 trials, a thermostatic device in another 5, an oral intervention in one trial, and topical dry eye medications in another. Additionally, one trial evaluated LipiFlow combined with an eyelid hygiene product versus eyelid hygiene products alone.

Five trials compared the efficacy of LipiFlow with the application of a basic warm compress, varying in duration and frequency. Only one trial included the addition of eyelid massage to the warm compress regimen. The analysis of symptom scores using the OSDI and the SPEED questionnaires revealed inconsistent results, showing no significant difference in symptoms between LipiFlow and warm compresses after 4 weeks. Furthermore, there was an absence of evidence indicating any significant difference in meibomian gland expression, meibum quality,

or TBUT when comparing LipiFlow to basic warm compresses. Similarly, another 5 trials contrasted LipiFlow with thermostatic devices. The analysis at 4 weeks revealed that thermostatic devices managed to reduce OSDI scores by a mean difference (MD) of 4.59 (95% confidence interval [CI], 1.23 to 7.95; $I^2=0$; $p=.007$; 553 participants; very low certainty evidence) compared to LipiFlow. Additionally, when LipiFlow combined with eyelid hygiene was compared to eyelid hygiene alone, no significant differences in signs or symptoms were observed at any evaluated time point. In a single trial, LipiFlow was compared against a topical dry eye disease medication, lifitegrast 5%. The trial suggested that lifitegrast 5% might enhance meibomian gland expression scores more effectively than LipiFlow by day 42 (MD, -1.21; 95% CI, -2.37 to -0.05; 50 participants; low certainty evidence), utilizing a MGES from 0 to 8. Another trial compared LipiFlow with an oral intervention, doxycycline, revealing that LipiFlow might significantly improve SPEED scores over doxycycline at 3 months (MD, -4.00; 95% CI, -7.33 to -0.67; 24 participants; very low certainty evidence). No other notable differences in signs or symptoms were observed between LipiFlow and doxycycline at 3 months. Additionally, no other statistically significant differences in symptoms or signs were identified in any other analyses conducted during this review within the 4-week timeframe. No trial reported any intervention-related, vision-threatening adverse events. LipiFlow shows comparable efficacy to other commonly used dry eye disease treatments in terms of signs and symptoms. However, the best level of evidence was deemed to have a high level of bias, resulting in low to very low certainty. Additional research with adequate masking, a standardized testing methodology, and a sample representative of the MGD population is needed before any definitive conclusions can be drawn regarding comparative benefits and harms of eyelid thermal pulsation therapy.

Tao et al. (2023) reported results of a systematic review that informed an 'Ophthalmic Technology Assessment' commissioned by the American Academy of Ophthalmology (AAO). (13) The review was designed to assess the efficacy and safety of thermal pulsation in improving signs or symptoms of MGD and dry eye compared with no therapy or conventional (nonthermal pulsation) therapy such as warm compress or eyelid hygiene. The literature search was performed in March 2023. For each study, the quality of study methodology was rated according to the AAO guidelines. Eight studies were rated as providing level I evidence (well-designed and well-conducted RCTs and systematic reviews) and 3 studies were rated as providing level II evidence (well-designed cohort studies and nonrandomized controlled cohort or follow-up trials). All included studies evaluated the LipiFlow device. The review did not include a meta-analysis. The authors stated that 9 (of 11) studies reported greater efficacy with LipiFlow compared to standard warm compress therapy and eyelid hygiene. In general, improvements were detected in both subjective and objective metrics of MGD within 1 to 12 months of thermal pulsation treatment compared with nontreatment. The authors noted that durability beyond several months is uncertain.

The RCTs included in these systematic reviews can be compared in Table 2.

Table 2. Comparison of RCTs Included in Systematic Reviews & Meta-Analyses

Study	Ben Ephraim Noyman et al. (2025) (11)	Pucker et al. (2024) ^a (12)	Tao et al. (2023) (13)
Baumann (2014)		●	
Blackie (2016)			●
Blackie (2018)			●
Booranapong (2020)	●	●	
Finis (2014)			●
Gupta (2022)	●	●	
Hagen (2018)		●	
He (2018)	●		
Holland (2022)	●	●	
Kasetsuwan (2020)	●	●	●
Lane (2012)	●		●
Li (2022)		●	
Matossian (2023)	●		
Mencucci (2023)	●	●	●
Meng (2023)	●	●	
Park (2021)	●		
Schlatter (2023)	●		
Tauber (2020a)	●	●	
Tauber (2020b)		●	
Wesley (2022)	●	●	
Zhao (2021)			●

RCT: randomized controlled trial.

^a Excluded from this table were 2 trials included in the Pucker (2024) systematic review as both trials were published in other languages (Baumann, 2014 in French; He, 2018 in Chinese).

Randomized Controlled Trials

Two RCTs of eyelid thermal pulsation for the treatment of dry eye syndrome have been published since publication of the above systematic reviews. Both trials are industry-sponsored studies.

Sadri et al. (2024) conducted a randomized (assessor-masked) trial (NCT05162261) to determine the efficacy and safety of thermo-mechanical action compared to LipiFlow in MGD. (14) Participants, recruited between 2022 and 2023 across 5 U.S. centers, who had OSDI scores between 23 and 79 and fluorescein TBUT of <10 seconds in each eye, were treated with either bilateral thermo-mechanical action (TMA) using the Tixel device (Novoxel) or thermal pulsation with LipiFlow. The TMA cohort underwent 3 treatment sessions 2 weeks apart, while the thermal pulsation group received a single session. Primary efficacy outcomes, including TBUT and OSDI, were assessed at baseline, at the 4-week mark, and 12 weeks post the final treatment session. Among the 106 participants (53 per group), TBUT showed significant improvements ($p < .001$), increasing by 3.0 ± 3.2 and 3.1 ± 4.3 seconds after TMA, and 2.7 ± 2.7

and 3.3 ± 3.6 seconds after thermal pulsation, at Week 4 and Week 12, respectively. Notably, the change in TBUT for TMA was proven to be non-inferior to thermal pulsation (linear mixed-effects model, $p < .001$). OSDI improved by 26.4 ± 21.1 and 28.6 ± 22.4 after TMA and 18.8 ± 21.0 and 21.9 ± 18.5 after thermal pulsation, at Week 4 and Week 12, respectively. No device-related adverse events occurred in either group.

Ayres et al. (2023) conducted a randomized (assessor-masked) controlled superiority trial (SAHARA, NCT04795752) to determine the efficacy and safety of TearCare (Sight Sciences) in comparison to topical cyclosporine 0.05% (CsA) for addressing dry eye disease in adults. (15) The trial enlisted 345 participants (172 in the TearCare group and 173 in the CsA group, recruited between 2021 and 2022) across 19 ophthalmic and optometric clinics in 11 U.S. states. Primary efficacy outcomes were changes from baseline in TBUT and OSDI at 6 months, with safety evaluations including adverse events, best corrected visual acuity, intraocular pressure, and slit-lamp observations. TBUT improved at all time points in both groups ($p < .0001$), with TearCare demonstrating a notably greater enhancement compared to CsA ($p = .0006$). The OSDI also exhibited significant improvement in both groups at all time points ($p < .0001$), though no significant differences were observed between the 2 treatment arms. Both therapies were largely safe and well-tolerated. Of the 19 treatment-emergent adverse events recorded in each group (constituting 11%), only 2 in the TearCare group (1%) and 8 in the CsA group (5%) were adjudged as related to the study treatment. All related adverse events were rated as mild ($n=9$) or moderate ($n=1$) in severity.

Ayres et al. (2024) published phase 2 results of the SAHARA study (a 6-month extension of SAHARA), which involved a single dose of TearCare in patients who had previously received CsA for 6 months ($n=322$ eyes). (16) After crossover to TearCare, TBUT improved to 6.6 ± 3.2 seconds at 3 months and 6.1 ± 2.8 seconds at 6 months (both $p < .001$ vs. baseline). There was no improvement in OSDI scores at 3 months, but there was a significant change in OSDI scores at 6 months ($p < .0001$ vs. baseline).

Hovanesian et al. (2025) published results of phase 3 of the SAHARA study (long-term follow-up). (17) A total of 166 patients were followed for 24 months. Retreatment with TearCare was allowed if TBUT decreased to within 2 seconds of baseline and if OSDI scores increased by 15 points from the prior visit; 32 patients required additional treatment at a median time of 8 months. Improvements from baseline in all efficacy parameters was maintained until 24 months in the remaining 134 patients.

Observational Studies

Four observational studies have assessed the long-term outcomes of subjects who underwent LipiFlow treatment. Greiner et al. (2013) (18) evaluated 18 (of 30) participants from a single site of the Lane (2012) RCT (cited in the Tao systematic review above), (19) observing that while several outcomes remained significantly improved from baseline, the improvements were less pronounced at 1 year compared to 1 month. Finis et al. (2014) monitored 26 patients 6 months post-treatment, noting sustained improvements in several outcome measures. (20) Greiner et al. (2016) study of 20 patients found that most outcomes remained significantly improved up to

3 years compared to baseline. (21) A retrospective cohort study by Hura et al. (2020) compared dry eye disease markers and meibomian gland imaging between patients who underwent LipiFlow treatment (n=30) and those who declined this therapy (n=13). (22) At 1 year, the treatment group showed sustained improvements in visible meibomian gland structure, TBUT, corneal staining, and meibomian gland evaluation scores over the control group. However, SPEED scores and tear osmolarity did not show sustained improvement 1-year post-therapy.

Summary of Evidence

For individuals who have dry eye symptoms consistent with meibomian gland dysfunction (MGD) who receive eyelid thermal pulsation, the evidence includes systematic reviews, randomized controlled trials (RCTs), and observational studies. A 2024 Cochrane meta-analysis evaluated the LipiFlow system's efficacy and safety for dry eye disease through 13 RCTs with 1155 participants. The findings showed that LipiFlow was comparable to other treatments like warm compresses, thermostatic devices, prescription eye drops, and doxycycline, with no notable differences in symptoms or signs. However, the evidence was deemed of low to very low certainty due to a high risk of bias. Similarly, another systematic review commissioned by the American Academy of Ophthalmology revealed that thermal pulsation with LipiFlow was more effective for MGD and dry eye than conventional therapies such as warm compresses or eyelid hygiene. However, the review also highlighted some limitations, particularly concerning the treatment's long-term durability. Another systematic review of multiple therapies for MGD found similar efficacy between thermal pulsation devices and comparator treatments. Since the publication of the systematic reviews, 2 industry-sponsored RCTs examining eyelid thermal pulsation for dry eye syndrome have been published. A randomized, assessor-masked trial comparing the efficacy and safety of LipiFlow versus thermo-mechanical action was conducted in participants with MGD across 5 U.S. centers. The study involved 106 participants with primary efficacy outcomes assessed at baseline, 4 weeks, and 12 weeks post-treatment. Results showed significant TBUT improvements in both groups, with thermo-mechanical action proving non-inferior to LipiFlow, and no device-related adverse events were reported. A second randomized, assessor-masked controlled superiority trial was conducted to compare the TearCare pulsation system with topical cyclosporine 0.05% (CsA) in 345 participants across 19 clinics in 11 U.S. states. The trial found significant tear break-up time improvements in both groups, with TearCare showing greater enhancement, and notable Ocular Surface Disease Index improvements without significant differences between treatments. Both therapies were safe, with mild to moderate treatment-related adverse events occurring in a small proportion of participants. Observational long-term follow-up found that efficacy was maintained in 134 patients at 24 months without need for retreatment, but 32 patients required additional treatment with TearCare. Observational studies on LipiFlow have shown sustained treatment effects for most outcomes up to 3 years. Additional RCTs are needed before any definitive conclusions can be drawn about the comparative benefits and risks of eyelid thermal pulsation therapy. These trials should include adequate masking, standardized testing methodologies (e.g., evidence-based controls), and longer follow-up periods. This will help ensure that the results are reliable and applicable to a broader population. The evidence is insufficient to determine that the technology results in an improvement in the net health outcome.

Practice Guidelines and Position Statements

American Academy of Ophthalmology

In 2018, the American Academy of Ophthalmology (AAO) updated preferred practice patterns guidelines on dry eye syndrome. (7) These guidelines list "In-office, physical heating and expression of the meibomian glands (including device-assisted therapies, such as LipiFlow, or intense pulse light treatment)" as one of several step-up treatments for patients who do not respond to conventional management, including the elimination of environmental factors and offending medications, dietary modifications, ocular lubricants, and lid hygiene and warm compresses.

In 2018, the AAO updated preferred practice patterns guidelines on blepharitis. (3) These guidelines cover the 3 clinical subcategories of blepharitis: staphylococcal, seborrheic, and meibomian gland dysfunction (posterior blepharitis specifically affects the meibomian glands). The following statements are made relevant to thermal pulsation treatment:

"There are also several in-office procedural treatments available that may theoretically unclog the inspissated meibomian gland orifices using intense pulsed light (IPL) or mechanical means (e.g., microblepharoexfoliation of the eyelid margin, meibomian gland probing, and/or devices using thermal pulsation). Although there have been industry-sponsored studies, independent, randomized, masked clinical trials have yet to be performed to assess efficacy of these costly, primarily fee-for-service treatments."

In 2023, the AAO updated preferred practice pattern guidelines on dry eye syndrome. These guidelines list thermal pulsation devices as a second-stage option for treatment of dry eye disease. (23) First-line treatment includes eyelid hygiene, warm compresses, and ocular lubrication. Other second-line treatment options include topical secretagogues, topical immunomodulators (such as cyclosporine), liftegrast, and short-term topical antibiotics and/or corticosteroids (for blepharitis).

In 2023, the AAO updated preferred practice pattern guidelines for blepharitis. (24) These guidelines indicate that multiple industry-sponsored studies have demonstrated that a single vectored thermal pulsation (VTP) treatment can be effective at improving meibomian gland function and reducing dry eye symptoms for a year or more post procedure. However, there have been no independent randomized controlled trials confirming or refuting these industry-sponsored studies.

"There are several in-office procedural treatments available that may improve the inspissated meibomian gland orifices using intense pulsed light (IPL) or theoretically unclog the meibomian glands by mechanical means (e.g., microblepharoexfoliation of the eyelid margin, meibomian gland probing, and/or devices using thermal pulsation). Although there have been industry-sponsored studies, independent, randomized clinical trials have yet to be performed to assess efficacy or superiority of any one of these treatments over another. [moderate quality, discretionary recommendation]"

Tear Film & Ocular Surface Society

The Tear Film & Ocular Surface Society is an industry-sponsored international organization that hosts periodic workshops on the topic of dry eye syndrome. (25) The management and therapy report of the latest dry eye workshop (DEWS) was published in 2025; the document includes a conflict of interest disclosure, but there is no description of the literature review methodology or system of evidence grading. For patients with a lipid-based dry eye etiology or with dry eye caused by meibomian gland dysfunction, the report concludes that the efficacy of lid heating devices is supported by randomized, placebo-controlled trials and are one of several treatment options. No recommendations are made for a preferred thermal pulsation device, but the report does comment on comparative efficacy. Regarding LipiFlow, non-industry sponsored studies have found that LipiFlow has similar efficacy as warm compresses and eyelid hygiene. Regarding iLux, the report states that it has statistically similar efficacy as LipiFlow. Regarding TearCare, a randomized study found similar efficacy between TearCare and LipiFlow overall, but patients with more severe symptoms did better with TearCare.

Ongoing and Unpublished Clinical Trials

Some currently ongoing or unpublished trials that might influence this policy are listed in Table 3.

Table 3. Summary of Key Trials

NCT Number	Trial Name	Planned Enrollment	Completion Date
NCT07365735 ^a	A Prospective, Randomized, Controlled Trial of the TearCare MGX™ System With and Without Warming Hold (Xtend)	20	May 2026
NCT06542276 ^a	Single Vectored Thermal Pulsation Treatment in Patients Using Topical Immunomodulators in the Management of Dry Eye Disease	30	Sep 2024
NCT05577910	Vectored Thermal Pulsation, Intense Pulsed Light, and Eyelid Warm Compress (VIEW) Therapies for Meibomian Gland Dysfunction- a Randomized, Assessor-masked, Active-controlled Clinical Trial	360	Jun 2025
NCT06483750	The Efficacy of the Systane iLux System on Preoperative Cataract Patients on Dry Eye Disease Due to Meibomian Gland Dysfunction	30	Feb 2025

NCT: national clinical trial.

^a Denotes industry-sponsored or cosponsored trial.

Coding

Procedure codes on Medical Policy documents are included **only** as a general reference tool for each policy. **They may not be all-inclusive.**

The presence or absence of procedure, service, supply, or device codes in a Medical Policy document has no relevance for determination of benefit coverage for members or reimbursement for providers. **Only the written coverage position in a Medical Policy should be used for such determinations.**

Benefit coverage determinations based on written Medical Policy coverage positions must include review of the member's benefit contract or Summary Plan Description (SPD) for defined coverage vs. non-coverage, benefit exclusions, and benefit limitations such as dollar or duration caps.

CPT Codes	0207T, 0330T, 0507T, 0563T
HCPCS Codes	None

*Current Procedural Terminology (CPT®) ©2025 American Medical Association: Chicago, IL.

References

1. Stapleton F, Alves M, Bunya VY, et al. TFOS DEWS II Epidemiology Report. Ocul Surf. Jul 2017; 15(3):334-365. PMID 28736337
2. Farrand KF, Fridman M, Stillman IÖ, et al. Prevalence of Diagnosed Dry Eye Disease in the United States Among Adults Aged 18 Years and Older. Am J Ophthalmol. Oct 2017; 182:90-98. PMID 28705660
3. Blepharitis. American Academy of Ophthalmology Cornea/External Disease Panel. Preferred Practice Pattern Guidelines. San Francisco, CA: American Academy of Ophthalmology; 2018. Available at <<https://www.aao.org>> (accessed April 15, 2026).
4. McCann P, Abraham AG, Mukhopadhyay A, et al. Prevalence and Incidence of Dry Eye and Meibomian Gland Dysfunction in the United States: A Systematic Review and Meta-analysis. JAMA Ophthalmol. Dec 01 2022; 140(12):1181-1192. PMID 36301551
5. Nichols KK, Foulks GN, Bron AJ, et al. The international workshop on meibomian gland dysfunction: executive summary. Invest Ophthalmol Vis Sci. Mar 30 2011; 52(4):1922-1929. PMID 21450913
6. Blackie CA, Korb DR, Knop E, et al. Nonobvious obstructive meibomian gland dysfunction. Cornea. Dec 2010; 29(12):1333-1345. PMID 20847669
7. Dry Eye Syndrome. American Academy of Ophthalmology Cornea/External Disease Panel. Preferred Practice Pattern Guidelines. San Francisco, CA: American Academy of Ophthalmology; 2018. Available at <<https://www.aao.org>> (accessed April 15, 2026).
8. Miller KL, Walt JG, Mink DR, et al. Minimal clinically important difference for the ocular surface disease index. Arch Ophthalmol. Jan 2010; 128(1):94-101. PMID 20065224
9. Ngo W, Situ P, Keir N, et al. Psychometric properties and validation of the Standard Patient Evaluation of Eye Dryness questionnaire. Cornea. Sep 2013; 32(9):1204-1210. PMID 23846405
10. Tomlinson A, Bron AJ, Korb DR, et al. The international workshop on meibomian gland dysfunction: report of the diagnosis subcommittee. Invest Ophthalmol Vis Sci. Mar 2011; 52(4):2006-2049. PMID 21450918
11. Ben Ephraim Noyman D, Chan CC, Teichman JC, et al. Dry Eye Disease Management Via Technological Methods: A Systematic Review and Network Meta-analysis. Ophthalmol Ther. Aug 2025; 14(8):1917-1954. PMID 40601205

12. Pucker AD, Yim TW, Rueff E, et al. LipiFlow for the treatment of dry eye disease. *Cochrane Database Syst Rev*. Feb 05 2024; 2(2):CD015448. PMID 38314898
13. Tao JP, Shen JF, Aakalu VK, et al. Thermal Pulsation in the Management of Meibomian Gland Dysfunction and Dry Eye: A Report by the American Academy of Ophthalmology. *Ophthalmology*. Dec 2023; 130(12):1336-1341. PMID 37642619
14. Sadri E, Verachtert A, Parkhurst GD, et al. Effectiveness and safety of a thermo-mechanical action device versus thermal pulsation device in the treatment of meibomian gland dysfunction. *J Cataract Refract Surg*. Dec 16 2024; 51(4):274-281. PMID 39680541
15. Ayres BD, Bloomenstein MR, Loh J, et al. A Randomized, Controlled Trial Comparing Tearcare® and Cyclosporine Ophthalmic Emulsion for the Treatment of Dry Eye Disease (SAHARA). *Clin Ophthalmol*. 2023; 17:3925-3940. PMID 38143559
16. Ayres BD, Bloomenstein MR, Loh J, et al. Improved Signs and Symptoms of Dry Eye Disease for Restasis® Patients Following a Single Tearcare® Treatment: Phase 2 of the SAHARA Study. *Clin Ophthalmol*. 2024; 18:1525-1534. PMID 38827771
17. Hovanesian J, Ayres BD, Bloomenstein MR, et al. Durability of the TearCare treatment effect in subjects with dry eye disease: Stage 3 of the Sahara randomized controlled trial. *Optom Vis Sci*. Aug 01 2025; 102(8):495-504. PMID 40719437
18. Greiner JV. Long-term (12-month) improvement in meibomian gland function and reduced dry eye symptoms with a single thermal pulsation treatment. *Clin Exp Ophthalmol*. Aug 2013; 41(6):524-530. PMID 23145471
19. Lane SS, DuBiner HB, Epstein RJ, et al. A new system, the LipiFlow, for the treatment of meibomian gland dysfunction. *Cornea*. Apr 2012; 31(4):396-404. PMID 22222996
20. Finis D, König C, Hayajneh J, et al. Six-month effects of a thermodynamic treatment for MGD and implications of meibomian gland atrophy. *Cornea*. Dec 2014; 33(12):1265-1270. PMID 25321941
21. Greiner JV. Long-Term (3 Year) Effects of a Single Thermal Pulsation System Treatment on Meibomian Gland Function and Dry Eye Symptoms. *Eye Contact Lens*. Mar 2016; 42(2):99-107. PMID 26222095
22. Hura AS, Epitropoulos AT, Czyz CN, et al. Visible Meibomian Gland Structure Increases After Vectored Thermal Pulsation Treatment in Dry Eye Disease Patients with Meibomian Gland Dysfunction. *Clin Ophthalmol*. 2020; 14:4287-4296. PMID 33324034
23. Amescua G, Ahmad S, Cheung AY, et al. Dry Eye Syndrome Preferred Practice Pattern®. *Ophthalmology*. Apr 2024; 131(4):P1-P49. PMID 38349301
24. Lin A, Ahmad S, Amescua G, et al. Blepharitis Preferred Practice Pattern®. *Ophthalmology*. Apr 2024; 131(4):P50-P86. PMID 38349296
25. Jones L, Craig JP, Markoulli M, et al. TFOS DEWS III: Management and Therapy. *Am J Ophthalmol*. Nov 2025; 279:289-386. PMID 40467022

Centers for Medicare and Medicaid Services (CMS)

The information contained in this section is for informational purposes only. HCSC makes no representation as to the accuracy of this information. It is not to be used for claims adjudication for HCSC Plans.

The Centers for Medicare and Medicaid Services (CMS) does not have a national Medicare coverage position. Coverage may be subject to local carrier discretion.

A national coverage position for Medicare may have been developed since this medical policy document was written. See Medicare's National Coverage at <<https://www.cms.hhs.gov>>.

Policy History/Revision	
Date	Description of Change
07/01/2026	Policy reactivated. Eyelid thermal pulsation therapy (e.g., LipiFlow® Thermal Pulsation System, TearCare® System) to treat dry eye syndrome is considered experimental, investigational and/or unproven. Policy name changed from Eyelid Thermal Pulsation.
12/31/2025	Document became inactive.
12/01/2024	Document updated with literature review. The following change was made in Coverage: Added the TearCare® System as an example of eyelid thermal pulsation therapy. Added references 4, 6, 7, 15-18, 23-28, 39; some updated, others removed.
08/15/2023	Reviewed. No changes.
12/01/2022	Document updated with literature review. Coverage unchanged. Added references 1-3, 11-14, 18, 23, 27-29.
08/01/2021	Reviewed. No changes.
09/01/2020	CPT codes updated.
07/15/2020	Document updated with literature review. The following changes were made to Coverage: 1) Separated statement on eyelid thermal pulsation therapy (e.g., LipiFlow® Thermal Pulsation System) and tear film imaging (e.g., LipiView II® Ocular Surface Interferometer) into two separate statements; and 2) Added statement on near infrared dual imaging (e.g., LipiScan™ Dynamic Meibomian Imager) as experimental, investigational and/or unproven. Added references 4-6, 10, 18-20, and 23. Title changed from “Eyelid Thermal Pulsation Therapy”.
07/15/2018	Document updated with literature review. Coverage unchanged. Added references 9, 13. Document title change from: Eyelid Thermal Pulsation Therapy for Dry Eye Syndrome.
07/15/2017	Reviewed. No changes.
07/15/2016	Document updated with literature review. Coverage unchanged.
07/01/2015	Reviewed. No changes.
07/15/2014	Document updated with literature review. Coverage unchanged.
11/01/2013	New medical document. Eyelid thermal pulsation therapy (which may include the use of the LipiView® for diagnosis and/or the LipiFlow® for treatment) is considered experimental, investigational, and unproven for all indications including but not limited to dry eye syndrome.